

Indian Institute of Technology Kanpur
Center for Lasers and Photonics
Department of Mechanical Engineering

August 27, 2013

Enquiry Number: CELP/PKP/INST/NC3,

Dated: August 27, 2013

Opening date: August 26, 2013 at 4:00 PM

Closing date: September 5th, 2013 at 5:00 PM

Sealed quotations are invited for the supply of picoammeter with the following specification:

Specifications:

Current range	20fA to 20 mA
Speed	1000 readings/sec
Source voltage	200 μ V – 500 V
Resolution	10fA (femtoAmper)

Features :

Source Mode	DC/AC
Voltage Sweep	Yes
Analog Input	Yes
Overload Protection	Yes

Terms and conditions:

- Technical and financial details should be in separate envelope.
- Maximum educational discounts should be applied
- Validity of quotation should be at least for 60 days.
- Price should be on FOR IIT Kanpur & should include the installation and training cost.
- Institute is exempted for payment of Excise Duty under notification No. 10/97.
- Warranty/Guarantee should be clearly mentioned.
- Normal payment terms for the institute will be applicable (90% on delivery of the items and remaining 10% after satisfactory installation/inspection).
- Quotation should carry proper certifications like agency certificate, proprietary certificate, etc.
- The delivery should be specifically stated. Earlier delivery may be preferred.
- The indenter reserves the right to withhold placement of final order. The right to reject all or any of the quotations and to split up the requirements or relax any or all of the above conditions without assigning any reason is reserved.

Kindly send the quotation (in duplicate) in sealed envelope to the following address.

Prof. P. K. Panigrahi
Department of Mechanical Engineering
Indian Institute of Technology, Kanpur, 208016
panig@iitk.ac.in