



Indian Institute of Technology, Kanpur

Department of Chemical Engineering

Dr. Yogesh M. Joshi

Associate Professor

Tel.: +91 512 2597993; Fax: +91 512 2590104

E-mail: srisiva@iitk.ac.in

December 27, 2011

Quotation Request Notice

Sealed quotations are invited from dealers/distributors by 'Water Purification System' of at least following specifications. The quote should be submitted to the Department of Chemical Engineering, IIT Kanpur by 29.12.2011.

Enquiry No: CHE/CF/2/2011

Opening date: 27.12.2011

Closing date: 06.01.2012

The Following Specifications are as under:

a) PRE FILTRATION STAGE: One stage purification step involving 5 micron and 1 micron filter.

b) ANALYTICAL GRADE WATER SYSTEM:

1. The system should respond favourably to feedwater having

- Fouling Index greater than 12
- Total Chlorine greater than 2ppm.
- Feedwater conductivity more than 2000 microS/cm

2. Four stage purification process:

- Primary purification by a pre-filter(with anti-scaling and activated carbon for bacteriostasis)
- Secondary purification through RO membrane to remove impurities with MW >200 daltons
- Third purification step should involve a self regenerating Electro deionization module with carbon bead at cathode to avoid de-ionization cartridge replacement and hazardous chemical regeneration.

3. Conductivity meter present before and after the RO cartridge to monitor RO cartridge efficiency.

4. Product water quality-

Resistivity	:>5Megohm-cm
TOC	:<30ppb
CONDUCTIVITY	:< 0.2uS/cm
Flow rate	: 3 L/hr

c) ULTRAPURE WATER SYSTEM

1. Purification process with ultrapure cartridge and an absolute 0.22 micron PVDF membrane final filter in stacked configuration.
2. The resistivity cell should be coaxial with a cell constant of 0.01cm^{-1} .
3. The system should have automatic re circulation, an alphanumeric backlit LCD display with auto-diagnostic features and alarms.
4. The system should have a provision to connect an Ultra Filtration cartridge to produce RNA's DNA,s free water.
5. The system should have a facility for volumetric dispensing as an option, with flow rate upto 1 litre/min.
6. Product water quality-

Resistivity	:18.2 Meg-ohm
Bacteria	:< 1cfu/ml
TOC	:<10 ppb
Particulates	:<1/ml
Pyrogen levels	:<0.001 EU/ml
Flow rate	:1 l/min

TANK specifications

1. A 50 Litre tank to store purified water
2. Should possess a sensor rod float switch for determining the level of stored water.

Dr.Yogesh M. Joshi

Associate Professor

Department of Chemical Engineering

Indian Institute of Technology Kanpur

Kanpur, Uttar Pradesh, PIN 208 016

India

Tel.: +91 512 2597697; Fax: +91 512 2590104

E-mail: srisiva@iitk.ac.in