

INDIAN INSTITUTE OF TECHNOLOGY, KANPUR
CHEMISTRY DEPARTMENT

Dated : 17/02/2012.

Enquiry No. CHM/PKB/2011-12/10

With below specs would be required for your Requirement for Digital Polarimeter

1. Optical rotation range ± 90 Deg
2. High precision (accuracy : at least ± 0.002 Deg Arc optical Rotation)
3. Prism : High quality
4. Wave length: 6 wave length - 365nm,405,436,546 ,589nm,633nm
5. Display : LCD
6. 3 years warranty

Sealed quotations have been invited by the undersigned on or before 23/02/2012.

Thanking you

With regards

(PK Bhardwaj)
Head Chemistry Department
IIT Kanpur