



Indian Institute of Technology Kanpur
Kanpur 208 016, India

Dr. Pratik Sen
Assistant Professor, Department of Chemistry
e-mail: psen@iitk.ac.in
Phone: +91-512-259-6312
Fax: +91-512-259-7436

Date: 22 August 2013

Enquiry No.: CHM/PS/2013-14/T-08

Last date: 02 September 2013

Sub.: Request for quotation for Chiller

Dear Sir,

Please send a sealed quotation with all the technical details of the following item(s) by 5:00 pm, 02 September 2013 to the following address.

Dr. Pratik Sen
Department of Chemistry
IIT Kanpur
Kanpur 208016
UP, India

Email: psen@iitk.ac.in

Thanking you,

Sincerely,

Pratik Sen
(Pratik Sen)

Item: **Recirculating Chiller -20 to 100 deg C**

Specifications:

- Capacity: 7 liter or more
- Temperature range -20 to 100 deg C or better
- Temperature stability: 0.1 deg C or better

Terms and Conditions:

1. Installation and demonstration at IIT Kanpur
2. Warranty: 1 year or more