

**INDIAN INSTITUTE OF TECHNOLOGY
CENTRE FOR ENVIRONMENTAL SCIENCE AND ENGINEERING**

Enquiry letter for purchase of High Volume Sampler for PM2.5

Sub: Quotation for supply, installation, commissioning and training of High Volume Sampler for PM 2.5 (Nos. 4)

Reference: IITK/CESE/2017-18/1010

Dated: July 28, 2017

Sir/Madam,

With reference to the subject mentioned above, you are invited to submit the quotation in a sealed cover. Configuration/Specifications are given below:

Specifications of High Volume Sampler for PM 2.5

Flow Controller	Volumetric Flow Controller
Motor Style	Brush style motor
PM2.5 Size Selective Inlet	PM2.5 Impactor Jets and Porous disc that is wetted with oil
Timer	USEPA recommended digital Timer with LCD screen, Internal Battery Backup, Default and Custom Sampling Schedules and Internal Data Log
Pressure Recorder:	Dickson 24-hour chart recorder
Elapsed Time Indicator:	Digital Multiple elapsed time indicator , Hours and Tenths
Flow Rate	40 CFM
Filter Holder:	Stainless Steel, 8" x 10"
Calibration Equipment	Variable Flow Calibration Kit

The quotation should have the following details:

1. Cost of the item and accessories and installation charges, if any
2. Technical specifications in detail
3. Technical bid and Price bid to be sealed separately (Two-bid system)
4. Warranty period
5. Educational discount considering end use for research and teaching
6. Payment terms
7. Proprietary Certificate, if applicable
8. Comprehensive AMC prices should be quoted separately
9. Any other relevant details

Terms and condition:-

1. Sealed Quotation must reach the undersigned on or **before August 10, 2017**.
2. Prices should be in USD and CIF Delhi.
3. Our Institute is partially exempted from custom duty.
4. The final selection will be made based on weights given to technical merit and pricing as 70% and 30% each, respectively.

Dr. Sachchida Nand Tripathi
Professor
Department of Civil Engineering
Indian Institute of Technology
Kanpur – 208016

Tel. 0512 – 2597845
Fax No. 0512-2597395
E-mail: snt@iitk.ac.in

