

Indian Institute of Technology - Kanpur
Department of Biological Sciences & Bioengineering

Enquiry Number: BSBE/IG/NC/SG3, dated: 19/09/2013

Sub.: Inquiry for the supply of laboratory work surface (clean, vibration free work bench)

Opening date: September 19, 2013 at 10:00 AM

Closing date: September 23, 2013 at 5:00 PM

Sealed quotes (technical bid and price bid separately sealed) are invited for the supply of laboratory work surface (laboratory tables) as per the specifications given in the next page.

Your quote should mention/include the following:

- Maximum discount if any should be offered and mentioned.
- Quoted price should include the cost for installation, warranty and required accessories (see below).
- Validity of the quote at least for 60 days.
- FOB (indicating port of shipment) and CIF (New Delhi) values should be quoted separately if import is required. For quotes in INR, the price quote should be for delivery at Kanpur.
- The quote should cover insurance for transport up to Kanpur.
- Indian agency commission if applicable (should be certified by the principal if no agency commission is applicable) in case of import.
- Authorization certificate from the principal if you are a local agent.
- Terms and conditions for the payment, including the banker's name of the principal and the account number, if any, for electronic transfer.
- Include proprietary item certificate if applicable.
- Technical literature to support your product (in technical bid).
- Users' list with contact address in technical bid.

Note: Companies submitting offers should have supplied similar products to biotechnology establishments and this should be supported by documents.

The quote should reach the undersigned on or before 5 pm on or before September 23, 2013. The envelope should be marked as "*Quote for the supply of laboratory work surface*".

The Head
Department of Biological Sciences & Bioengineering
Indian Institute of Technology,
Kanpur 208016 (UP)

Inquiry for the supply of laboratory work surface (laboratory table)

Laboratory table for molecular biology lab (Quantity 4)

Specification:

Laboratory table with modular features as specified below (specification for one set):

- Total work surface for one table: 150"x60"x36" dimension (W/D/H)
- The above table should have 5 independent modular components, each having 60"x30" (W/D), onsite assembly of these five should result in the table of the dimension 150"x60"x36".
- The load bearing frame of table should be made up of at least 30mm square pipe (good quality iron), with level adjuster and black color paint finish. The assembled table should be able to house heavy load machines as table-top items.
- The design of the table should be that the front area of the table is free of support shafts therefore cabinets or equipments can be housed underneath.
- The table top should have black color granite top, supported below by minimum 12 mm ISI quality heat pressed ply board.
- The table should have two sample racks, each with a dimension of 60"(W)x12"(D)x36"(H), having three shelves, separated by 12", and 4 15/5 amp combo power points, switches and an overload safety switch. The two racks should be designed that they can be mounted on the assembled table.

Warranty period: Minimum 1 year from the date of installation