

Indian Institute of Technology Kanpur
Department of Civil Engineering

28 August, 2012

Sub: Quotation for Digital Sieve Shaker

Dear Sir,

We propose to purchase a digital sieve shaker for our lab. You are requested to give quotation with best price/discounts with following preferable specifications.

Grain size (diameter)		ASTM
ϕ (phi)	mm/ μ m	(U.S. Standard)
-2	4 mm	5
-1	2	10
0	1	18
1	500 μ m	35
2	250	60
	212	70
	150	100
3	125	120
	90	170
4	63	230
	45	325
5	31	450
6	25-16	500

Sl. No.	Mesh Size (Microns)	ASTM-E11 Sieve no	Diameter
1	250 μ m	60	< 4 "
2	180 μ m	80	< 4 "
3	125 μ m	120	< 4 "
4	63 μ m	230	< 4 "

You are requested to submit/post your quotation in a sealed envelope with complete product description, technical literature, price, warranty period, delivery time and other terms and conditions by September 10, 2012. If you have any question please call Dr. J. N. Malik at 2597723, email: javed@iitk.ac.in.

Yours Sincerely



Dr. J. N. Malik
In charge EGS Lab