



Enquiry No. : CHE/GD/BRNS/2012/01

January 27, 2012

Quotations are invited from interested suppliers for **Mass Flow Controllers** for measurement of Oxygen and Argon in a continuous flow system as per the following specifications.

Specifications

Measuring Gas Flow: Oxygen, Argon

Flow Range: 0.02-1.0 LPM (Oxygen), 0.04-2.0 LPM (Argon)

Inlet Pressure: 2 bar

Outlet Pressure: 1 bar

Operating temperature: Room temperature (25⁰C)

End connection: 1/4" Swagelok preferred

Resolution: 0.001 ml/min

Kindly mention about the terms and conditions such as;

- * Validity of the quotation for a minimum of 60 days.
- * Warranty period
- * Mention the FOB rate.
- * Special educational discounts, if any.

For indigenous or import items please use the form number 603 or 608 respectively given at

<http://web.iitk.ac.in/dord/rndforms/603.pdf>

<http://web.iitk.ac.in/dord/rndforms/608.pdf>

The quotations are to be sent to the under signed by February 07, 2012.

Thanking you;

Goutam Deo

Professor, Department of Chemical Engineering

CL 201B, IIT Kanpur

Kanpur- 208 016

Ph: 0512-2597881, Fax: 0512-2590104