

# Indian Institute of Technology Kanpur

Dr. Ashis K. Patra, Ph.D.  
Assistant Professor



Department of Chemistry  
Core Laboratory, Room No. 201E  
IIT Kanpur, Kanpur 208016, India  
Tel: +91-512-259-6780  
FAX: +91-512-259-7436  
E-mail: akpatra@iitk.ac.in

Date: August 26, 2013

**Enquiry No: CHM/AKP/2013-14/02**

**Closing Date: September 08, 2013**

**Subject: Request for quotation for Vertical Autoclave**

Dear Sir/Madam:

Sealed quotations are invited **very urgently** for purchase of **Vertical Autoclave** for laboratory use for sterilization purpose with following minimum specifications:

- Double/tripled walled stainless steel vertical construction of chamber and lid and ergonomic design with stand and having autoclavable basket.
- Double safety radial locking mechanism, double safety valve protection.
- Stainless steel steam release valve, equipped with pressure gauge.
- Standard working temperature 120 °C, working pressure range 5-20 psi.
- Foot paddle lift for lid opening mechanism for ease of operation.
- Fitted with timer with alarm system for auto cut-off the current from main.
- Fitted with digital temperature indicator or programmable temperature controller.
- Capacity around 75-100 liters.

**Terms & Conditions:**

1. Prices should be in INR, inclusive all taxes and CIF to IIT Kanpur only and include the installation.
2. Warranty must be properly maintained in your quotation.
3. Validity of quotations should be at least for 15 days. Delivery within 2 weeks.
4. Our organization is an educational institute and liable to get maximum educational discount from manufacturer. IITK is exempted from excise/custom duty.

Sealed envelope carrying best offer quotation and printed literature with technical specification should reach undersign by **8<sup>th</sup> September 2013, 5:00 PM** to the following address **preferably by speed post/courier**.

**Dr. Ashis K. Patra,**  
**Department of Chemistry,**  
**Core Lab. 201E,**  
**IIT Kanpur, Kanpur-208016, U.P., INDIA**  
**E-mail: akpatra@iitk.ac.in**