



Indian Institute of Technology Kanpur
Kanpur 208 016, India

Dr. Pratik Sen
Assistant Professor, Department of Chemistry
e-mail: psen@iitk.ac.in
Phone: +91-512-259-6312
Fax: +91-512-259-7436

Date: 22 August 2013

Enquiry No.: CHM/PS/2013-14/T-02

Last date: 02 September 2013

Sub.: Request for quotation for Refractometer

Dear Sir,

Please send a sealed quotation with all the technical details of the following item(s) by 5:00 pm, 02 September 2013 to the following address.

Dr. Pratik Sen
Department of Chemistry
IIT Kanpur
Kanpur 208016
UP, India
Email: psen@iitk.ac.in

Thanking you,

Sincerely,

Pratik Sen
(Pratik Sen)

Item: **Digital Refractometer (Abbe)**

Specifications:

1. RI-range: 1.3000 to 1.7000 or better
2. Readability: 0.0001
3. Accuracy: 0.0002 or better
4. Measurement temperature range: 5 to 50 deg C with temperature accuracy of 0.1 deg C or better
5. Required light source, power source, etc. to run the system
6. Installation and demonstration at IIT Kanpur
7. Warranty: 1 year or more

Quantity: 1

Terms and Conditions:

- (a) A single company / distributor has to quote all the required items for stand alone operations
- (b) Delivery within ~month